		F HEALTH AND HUMAN SERVICES ID DRUG ADMINISTRATION	5		
DISTRICT OFFICE ADDRESS AND PHONE NUMBER 10 Waterview Blvd, third floor Parsippany, New Jersey, 07054			DATE(S) OF INSPECTION 08/27-30/2018, 09/13/2018, 09/25/2018 & 09/27/2018 FEI NUMBER		
973-331-4900			3011761882		
Industry Information: www.fda.gov/oc/industry NAME AND TITLE OF INDIVIDUAL TO WHOM REPORT IS ISSUED					
NAME AND TITLE OF INDIV	IDUAL TO WHOM REPORT IS ISSUED				
TO: Mr. Mark Taylor	, RPh, FACVP, Chief Executive Office	er			
FIRM NAME		STREET ADDRESS	CONTROL OF THE CONTRO		
Curexa 3007 C		3007 Ocean Heights Av	e.		
CITY, STATE AND ZIP CODE		TYPE OF ESTABLISHMENT IN	TYPE OF ESTABLISHMENT INSPECTED		
Egg Harbor, New Jersey 08234		Producer of sterile and a	Producer of sterile and non-sterile Prescription drugs		
THIS DOCUMENT LISTS OBSERVATIONS MADE BY THE FDA REPRESENTATIVE(S) DURING THE INSPECTION OF YOUR FACILITY. THEY ARE INSPECTIONAL OBSERVATIONS; AND DO NOT REPRESENT A FINAL AGENCY DETERMINATION REGARDING YOUR COMPLIANCE. IF YOU HAVE AN OBJECTION REGARDING AN OBSERVATION, OR HAVE IMPLEMENTED, OR PLAN TO IMPLEMENT CORRECTIVE ACTION IN RESPONSE TO AN OBSERVATION, YOU MAY DISCUSS THE OBJECTION OR ACTION WITH THE FDA REPRESENTATIVE(S) DURING THE INSPECTION OR SUBMIT THIS INFORMATION TO FDA AT THE ADDRESS ABOVE. IF YOU HAVE ANY QUESTIONS, PLEASE CONTACT FDA AT THE PHONE NUMBER AND ADDRESS ABOVE. DURING AN INSPECTION OF YOUR FIRM (I) (WE) OBSERVED:					
Sterile, injectable liquid dosage form preparation					
OBSERVATIONS 1					
Procedures designed to prevent microbiological contamination of drug products purporting to be sterile are not					
established, written or followed.					
Specifically,					
Records to demonstrate the cleanroom meets the required grade suitable for sterile filling are inadequate.					
For example.					
a. Air changes reported by the contracted provider of service do not meet ISO 5 and ISO 6 standards. b. Although the suite designated for sterile processing uses magnehelic gauges, there are no records to demonstrate the differential pressure conditions suitable for ISO 5, ISO 6 and ISO7 are maintained during sterile filling. c. Environmental monitoring records do not include continual sampling of non-viable particles to show the rooms/hoods classified as ISO 5, ISO 6 or ISO 7 maintain acceptable conditions for intended operations.					
Inspection dates: 08/27/2018, 08/28/2018, 08/29/2018, 08/30/2018, 09/13/2018, 09/25/2018 & 09/27/2018					
EMPLOYE	EE(S) SIGNATURE	EMPLOYEE(S) NAME AND TITLE (PA	rint or Type)	DATE ISSUED	
SEE REVERSE OF THIS PAGE	WIA	Frederick Razzaghi, Investigato		09/27/2018	

FORM FDA 483 (9/08) PRÉVIOUS EDITION OBSOLETE